
THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

| REVISION | | | |
|---|---|---|---|
| REV. | DESCRIPTION | DATE | APPROVED |
| B | RoHS Compliancy | 06-01-06 | D.P. |
| | | | |
| | | | |




**CONTROLLED**

NOTES:

1. CORDAGE: SJT, 18AWG / 3C, UNSHIELDED, NA COLOR CODE, TEMP. RATING 60°C, RATING: 125V 10A, JACKET: COLOR GRAY
2. PLUG: NEMA 5-15P HOSPITAL GRADE (CLEAR)
3. OTHER END: ROJ 2″, STRIP 5/8″
4. APPROVALS: UL, CSA
5. ROHS COMPLIANT
6. CAUTION LABEL

| Dim range (mm) | Tolerance +/- (mm) |
|---|---|
| 0-1 | 0.15 |
| 1-6 | 0.20 |
| 6-18 | 0.30 |
| 18-50 | 0.50 |
| 50-120 | 0.70 |
| 120-250 | 0.90 |
| 250-500 | 1.20 |

# Qualtek Electronics Corp.
## PPC DIVISION

Part Number: 232002-06 RoHS Compliant

Description: HOSPITAL GRADE POWER CORD

| | UNIT: in | REV. B |
|---|---|---|
| Drawn / Date | Checked / Date | Approved / Date |
| B. ZHANG 01-21-04 | 01-21-04 | JJH / 01-21-04 |