


THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

REVISION			
REV.	DESCRIPTION	DATE	APPROVED
B	RoHS Compliancy	06-01-06	DP



NOTES:

- 1. CORDAGE: SJT, 18AWG / 3C, UNSHIELDED, NA COLOR CODE, TEMP. RATING 60°C, RATING: 125V 10A, JACKET COLOR: GRAY
- 2. PLUG: NEMA 5-15P HOSPITAL GRADE (GRAY)
- 3. OTHER END: ROJ 2"STRIP 5/8"
- 4. APPROVALS: UL,CSA
- 5. ROHS COMPLIANT
- 6. CAUTION LABEL

CONTROLLED

Dim range (mm)	Tolerance +/- (mm)	Qualtek Electronics Corp. PPC DIVISION		
		Part Number:	242003-06	RoHS Compliant
0-1	0.15	Description: HOSPITAL GRADE POWER CORD		
1-6	0.20			
6-18	0.30	UNIT: in	REV.B	Approved / Date
18-50	0.50			
50-120	0.70	Drawn / Date	Checked / Date	Approved / Date
120-250	0.90	B.ZHANG 05-17-04	05-17-04	JJH / 05-17-04
250-500	1.20			